

MutaPLEX® Coronavirus real time RT-PCR kit (Immundiagnostik AG, cat. No. KG1926xx)

MutaPLEX® RespiraScreen 1 real time RT-PCR kit (Immundiagnostik AG, cat No. KG1927xx)

MutaPLEX® RespiraScreen 2 DIFF real time RT-PCR kit (Immundiagnostik AG, cat No. KG1928xx)

MutaPLEX® Coronavirus 4G real time RT-PCR kit (Immundiagnostik AG, cat. No. KG1936xx)

UPDATE 2022/06/02

Influence of mutations in the SARS-CoV-2 Spike protein gene on the detection:

In silico analysis by Immundiagnostik AG, analyzing the following spike protein gene mutations was performed on the available sequences in the global databases of gene bank (NCBI) and GISAID (EpiCov™).

The screened variants include the Variants of Concern / Variants of Interest / Variants under Monitoring:

VOC: **[Alpha]*, [Beta]*, [Gamma]*, [Delta]** (all AY Subvariants), **[Omicron]** (BA.1, BA.2, BA.4, BA.5)

VOI: **[Epsilon]*, [Zeta]*, [Eta]*, [Theta]*, [Iota]*, [Kappa]*, [Lambda]*, [Mu]***

VUM: **[Omicron]** (BA.3, BA.2 + L452X),

*De-escalated variants by at least one of the following criteria: (1) the variant is no longer circulating, (2) the variant has been circulating for a long time without any impact on the overall epidemiological situation, (3) scientific evidence demonstrates that the variant is not associated with any concerning properties. (ecdc.europa.eu/en/covid-19/variants-concern)

None of the listed mutations shows an impact on the sensitivity and specificity of the named real time RT-PCRs.

Place and Date of Release:

Bensheim, 02.06.2022

Head of Molecular Biology:



Dr. Wolfgang Reichert